

Ossur rebound air walking boot (short and long versions)

Indications

- Grade 2 & 3 sprain
- Stable Fractures
- Post-operative use
- Trauma
- Rehabilitation

Fitting instructions

- Once seated unwrap the Velcro,
- Put foot into the boot ensuring the heel is placed firmly into the bottom corner,
- Place the faceplate **INSIDE** the back shell ensuring that both the buttons (inflate and deflate) are inside the correct holes,
- Fasten up all the Velcro straps to make the boot supportive and comfortable.

Inflation/deflation

- Press the big button to push air into the air bag which is located around the back and sides of the leg.
- Inflate until the boot feels supportive. **DO NOT OVER INFLATE.**
- **Before** removal of the boot, press and hold the deflate button until all the air is gone



Care instructions

- Remove liner from the air-walker before washing.
- The liner should be washed by hand using a mild detergent.
- Do not wash in washing machine or place in dryer.
- Line dry in shade.
- Walker plastic (the main shell) can be washed with a damp cloth and mild soap.

List of materials

Liner: Polyester, Nylon, Polyurethane, PVC (no bpa), ABS

Shell: Polypropylene, EVA, TPO, Steel, Nylon, Cyanoacrylate

If you have any concerns or queries regarding your boot, please do not hesitate to contact us on the number below.

Contact details

Plaster room

Telephone: 0300 019 2140

Please note: our telephone lines are open from 9am to 5pm however a 24 hour answerphone service is available and we will get back to you as soon as we can

Poole Hospital, Longfleet Road, Poole, Dorset, BH15 2JB

Author: **Gary Smith** Date: **December 2021** Version: **Four** Review date: **December 2024** Ref: **594/21**

t: 01202 665511 w: www.uhd.nhs.uk : @UHD_NHS : @UHDTrust : @uhd_nhs